



**WorkStation Industries**

(o) 714-258-7535

(f) 714-258-1057

(e) sales@workstationindustries.com

(w) www.workstationindustries.com

### **Base cabinet Construction and Specifications**

- 100 % 18ga CRS heavy duty welded steel
- 4" toe kick
- xx W x 2D x 30 or 25h
- Removable rear panel using self-tapping sheet metal screws
- Levelers come with 12-gauge gusset leveling plate
- Glides are 100% Zinc plated and are adjustable inside the cabinet
- Welded front stringer or front and rear beam are made of 16-gauge steel
- and are fully welded to help make the cabinet square
- Powder coated paint has been positive tested at 600-hour salt test
- 5 yr limited warranty
- When tops are added it adds  $\frac{3}{4}$ - 1" to the overall height
- All cabinets have been tested and comply with SEFA 8 standards

### **Cabinet Drawer Construction and Specifications**

- 100lb 18" full extension ball bearing Drawer slide extensions
- Double wall drawer and door faces
- Full finger pulls, 4" lab grade aluminum pulls, or 4" aluminum cup finger pulls are all available
- Drawer bottoms are fully welded to prevent fluid from leaking
- Side mounted drawer slide brackets are welded to the side of the drawers to prevent contamination
- Powder coated paint has been positive tested at 600-hour salt test
- 5 yr limited warranty, 1yr for drawer slides
- Rubber bumpers are used for quite close
- $\frac{3}{4}$ " thick phenolic resin used as a back panel to counter balance the cabinet when drawers are pulled out